



Short Communication

Fast SARS-CoV-2 detection by RT-qPCR in preheated nasopharyngeal swab samples



Julia Alcoba-Florez^a, Rafaela González-Montelongo^b, Antonio Íñigo-Campos^b,
Diego García-Martínez de Artola^a, Helena Gil-Campesino^a,
The Microbiology Technical Support Team^{a,c}, Laura Ciuffreda^d,
Agustín Valenzuela-Fernández^{e,f}, Carlos Flores^{b,d,g,h,*}

^a Servicio de Microbiología, Hospital Universitario N. S. de Candelaria, 38010 Santa Cruz de Tenerife, Spain

^b Genomics Division, Instituto Tecnológico y de Energías Renovables, 38600 Santa Cruz de Tenerife, Spain

^c The members of the team are listed at the end of the article

^d Research Unit, Hospital Universitario N. S. de Candelaria, 38010 Santa Cruz de Tenerife, Spain

^e Laboratorio de Inmunología Celular y Viral, Unidad de Farmacología, Facultad de Medicina, Universidad de La Laguna, 38200 San Cristóbal de La Laguna, Spain

^f Red española de Investigación en VIH/SIDA (RIS)-RETIC, Instituto de Salud Carlos III, 28029 Madrid, Spain

^g CIBER de Enfermedades Respiratorias, Instituto de Salud Carlos III, 28029 Madrid, Spain

^h Instituto de Tecnologías Biomédicas (ITB) Universidad de La Laguna, 38200 San Cristóbal de La Laguna, Spain

ARTICLE INFO

Article history:

Received 4 May 2020

Received in revised form 21 May 2020

Accepted 24 May 2020

Keywords:

COVID-19

SARS-CoV-2

diagnosis

sample treatment

RNA extraction

fast protocols

ABSTRACT

Objectives: The gold-standard COVID-19 diagnosis relies on detecting SARS-CoV-2 using RNA purification and one-step retrotranscription and quantitative PCR (RT-qPCR). Based on the urgent need for high-throughput screening, we tested the performance of three alternative, simple and affordable protocols to rapidly detect SARS-CoV-2, bypassing the long and tedious RNA extraction step and reducing the time to viral detection.

Methods: We evaluated three methods based on direct nasopharyngeal swab viral transmission medium (VTM) heating before the RT-qPCR: a) direct without additives; b) in a formamide-EDTA (FAE) buffer, c) in a RNAsnap™ buffer.

Results: Although with a delay in cycle threshold compared to the gold-standard, we found consistent results in nasopharyngeal swab samples that were subject to a direct 70°C incubation for 10 min.

Conclusions: Our findings provide valuable options to overcome any supply chain issue and help to increase the throughput of diagnostic tests, thereby complementing standard diagnosis.

© 2020 The Author(s). Published by Elsevier Ltd on behalf of International Society for Infectious Diseases.

This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

Introduction

The ongoing coronavirus disease 2019 (COVID-19) pandemic due to Severe Acute Respiratory Syndrome Coronavirus 2 (SARS-CoV-2) worldwide infection (<https://www.who.int/emergencies/diseases/novel-coronavirus-2019/situation-reports>) has imposed

an unexpected high burden on the health care systems worldwide leading to an increasing demand for daily diagnostic screening. The current standard assay for diagnosis is based on the extraction of RNA from respiratory samples, especially from nasopharyngeal swab viral transport media (VTM), and subsequent one-step reverse transcription and real-time quantitative PCR (RT-qPCR) targeting one or several sequences from SARS-CoV-2 (Corman et al. 2020). However, this standard procedure usually takes 3.5–4.0 h considering the manual interventions, and there is a risk of reagent shortage in major kit suppliers, particularly for the RNA extraction step. Alternatives to accelerate this procedure have been proposed in consequence, the most efficient relying on Loop-mediated Isothermal Amplification (LAMP) (Esbin et al. 2020).

Here we aimed to simplify the current diagnostic standard for COVID-19 by skipping the RNA extraction step. We tested three

* Corresponding author at: Unidad de Investigación, Hospital Universitario N.S. de Candelaria, Carretera del Rosario s/n, 38010 Santa Cruz de Tenerife, Spain.

E-mail addresses: juliaalcoba@gmail.com (J. Alcoba-Florez), rgonzalezmontelongo@iter.es (R. González-Montelongo), ainigo@iter.es (A. Íñigo-Campos), diegogarciamartinezdeartola@gmail.com (D.G.-M. de Artola), helegc@hotmail.com (H. Gil-Campesino), laura.ciuffreda1988@gmail.com (L. Ciuffreda), avalenzu@ull.edu.es (A. Valenzuela-Fernández), cflores@ull.edu.es (C. Flores).

simple approaches based on direct nasopharyngeal swab VTM heating before the RT-qPCR: a) directly without additives (Direct); b) in a formamide-EDTA (FAE) buffer (Shedlovskiy et al. 2017); and c) in a RNAsnap™ buffer (Stead et al. 2012).

Materials and Methods

The study was conducted at the University Hospital Nuestra Señora de Candelaria (Santa Cruz de Tenerife, Spain) during March 2020. For the exploratory stage, we selected nasopharyngeal swabs from four COVID-19/SARS-CoV-2 patients and four COVID-19 negative controls. For the validation stage, 90 independent samples (41 COVID-19 positives and 49 negatives) were subjected to the treatment providing the smallest cycle threshold deviations from the standard protocol in the exploratory stage. Sample manipulation and diagnosis, and alternative protocols are detailed in Supplementary materials.

Results

Exploratory stage

The non-template control did not show amplification in any of the protocols both for the SARS-CoV-2 or the internal control (Table S1). The positive control for the E-gene amplification yielded positive results in the RT-qPCR experiments of the three alternative protocols. Furthermore, all samples gave positive results for the internal control. When RT-qPCR was carried out on the same four positive samples treated using the alternative protocols (FAE, RNAsnap™ and Direct), we observed amplification of the E-gene in all three conditions, although with a displacement of the Ct values (Table 1). Compared to the standard RNA extraction, we observed an average (\pm SD) increase in the Ct of 6.9 (\pm 1.7), 7.8 (\pm 1.7), and 8.5 (\pm 1.1) for the Direct, RNAsnap™ and FAE treatments, respectively (Figure S1).

Validation stage

Based on these results, we assayed 90 independent VTM samples from 41 COVID-19 positives (Table S2) and 49 negatives using the Direct method. We verified that all samples gave positive results for the internal control (average Ct of 29.6 \pm 2.5) although the amplification Ct was, on average, slightly larger than that obtained by the standard RNA extraction method in the same samples (average Ct of 27.0 \pm 1.5).

Out of the 41 COVID-19/SARS-CoV-2 positive VTM samples, only five did not yield amplification for the E-gene with the Direct treatment. Regarding the internal control results on the extracted

RNA of these five samples, we did not observe significant differences when compared with those from the other COVID-19/SARS-CoV-2 positive samples (average Ct of 27.6 \pm 1.2 and 26.9 \pm 1.6, respectively; p = 0.457). However, their Ct values for the E-gene were larger (average Ct of 34.0 \pm 2.0 and 25.7 \pm 4.9, respectively; p = 0.0007). Therefore, we considered these five samples as false negatives, corresponding to a false negative rate in the Direct treatment of 12% (95% confidence interval [CI] = 5–28). Considering the 36 samples that were COVID-19/SARS-CoV-2 positive by the two methods, there was an average increase in the E-gene Ct by the Direct method of 6.1 (\pm 1.6) compared to that obtained by a standard RNA extraction. None of the COVID-19/SARS-CoV-2 negative VTM samples was classified as positive by the Direct treatment. Therefore, the Direct method yielded a sensitivity, specificity and accuracy of 87.8% (95% CI = 73.8–95.9), 100% (95% CI = 92.8–100), and 99.9% (95% CI = 95.7–100), respectively.

Discussion

While the three heating treatments of the sample and direct use in the subsequent detection showed positive amplification of the SARS-CoV-2 E-gene, the Direct method provided the best results and were highly consistent with the COVID-19/SARS-CoV-2 infection diagnosis based on the standard RNA extraction (Fig. 1a) in nearly half of the time (Fig. 1b). We caution that the study was done with a limited number of samples and amplifications should be closely monitored to avoid increasing the false negatives above that of the standard diagnosis based on RNA extractions (Xie et al. 2020). Despite that, diverse empirical assessments of our protocol and that proposed by Fomsgaard & Rosenstjerne (2020) revealed that the quantitative results are highly comparable (Calvez et al. 2020). Remarkably, SARS-CoV-2 and SARS-CoV-1 show comparable environmental stability (van Doremalen et al. 2020), and several evidences suggest that SARS-CoV-1 (Geller et al. 2012) and SARS-CoV-2 (Pastorino et al. 2020) lose infectivity above 56 °C within short periods of time, and without any significant effect on the number of viral gene-copies detected by RT-qPCR below 92 °C, even after 30–60 min of pre-treatment (Pastorino et al. 2020). Therefore, we postulate that the Direct protocol at 70 °C for 10 min may also help to diminish the infectiveness of the samples, without significant viral RNA degradation during manipulation.

Finally, we warn that the choice of RT-qPCR kits might have impact on the sensitivity of the Direct protocol. As an example, the average increase in the Ct by the Direct method compared to the standard RNA extraction was 3.5 (\pm 2.0) using the newly released TaqPath COVID-19 CE-IVD RT-PCR Kit using their ORF1ab assay (Thermo Fisher Scientific).

Authors' contributions

JAF, RGM and CF designed the study. JAF, RGM, AIC, DGM, HGC, TMTST and CF participated in data acquisition. JAF, RGM and CF performed the analyses and data interpretation. LC, AVF, RGM and CF wrote the draft of the manuscript. All authors contributed in the critical revision and final approval of the manuscript.

The Microbiology Technical Support Team

M. Sonia Batista-Torres, Concepcio'n Beltra'n-Tacoronte, Esther G. Go'mez-Rui'z, Victoria Gonza'lez-Gonza'lez, Teodora Manrique-Izquierdo, Rocío M. Rivera-Ruiz, Iru Trujillo-Medina, Guacimara Espinel- Guerra, Chaxiraxi Medina-Coello, M. Candelaria Padilla-Marti'n.

Table 1
Main RT-qPCR results for SARS-CoV-2 E-gene amplification.

Sample	Diagnosis	Threshold cycle			
		RNA extraction	Direct	RNAsnap	FAE ^a
1	Positive	19.3	28.0	28.9	27.1
2	Positive	19.7	27.5	27.0	29.7
3	Positive	29.0	34.0	34.7	37.6
4	Positive	31.0	37.0	39.7	38.6
5	Negative	NA	NA	NA	NA
6	Negative	NA	NA	NA	NA
7	Negative	NA	NA	NA	NA
8	Negative	NA	NA	NA	NA
Non-template	-	NA	NA	NA	NA
Positive control	-	30.0	29.2	29.2	32.6

^a FAE, formamide-EDTA; NA, not available.




Fig. 1. Schematic diagram of the standard RNA extraction-based protocol (a) compared to the alternative fast protocol (b) and the timesaving estimates for RT-qPCR-based SARS-CoV-2 testing. Indicated times should be taken as conservative estimates as they will be dependent on the personnel skills and the number of samples being assessed on the experiment. Ct, cycle threshold.

Conflicts of Interest

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

Funding

This research was funded by Cabildo Insular de Tenerife, [grant number CGIEU0000219140]; the agreement with Instituto Tecnológico y de Energías Renovables (ITER) to strengthen scientific and technological education, training research, development and innovation in Genomics, Personalized Medicine and Biotechnology, [grant number OA17/008]; Ministerio de Innovación y Ciencia, [grant number RTI2018-093747-B-100 and RTC-2017-6471-1], co-funded by the European Regional Development Fund (ERDF); Lab P2+ facility, [grant number UNLL10-3E-783], co-funded by the ERDF and “Fundación CajaCanarias”; and the Spanish HIV/AIDS Research Network, [grant number RIS-RETIC, RD16/0025/0011], co-funded by Instituto de Salud Carlos III and by the ERDF.

Ethical Approval

The University Hospital Nuestra Señora de Candelaria (Santa Cruz de Tenerife, Spain) review board approved the study (ethics approval number: CHUNSC_2020_24).

Declaration of interests

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

Acknowledgments

We deeply acknowledge the University Hospital Nuestra Señora de Candelaria board of directors and the executive team for their

strong support and assistance in accessing diverse resources used in the study.

Appendix A. Supplementary data

Supplementary material related to this article can be found, in the online version, at doi:<https://doi.org/10.1016/j.ijid.2020.05.099>.

References

- Calvez R, Taylor A, Calvo-Bado L, Fink C. Molecular detection of SARS-CoV-2 using a reagent-free approach. *MedRxiv* 2020;(May), doi:<http://dx.doi.org/10.1101/2020.04.28.20083626>.
- Corman VM, Landt O, Kaiser M, Molenkamp R, Meijer A, Chu DKW, et al. Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR. *Euro Surveill* 2020;25 (Jan (3)):2000045.
- van Doremalen N, Bushmaker T, Morris DH, Holbrook MG, Gamble A, Williamson BN, et al. Aerosol and Surface Stability of SARS-CoV-2 as Compared with SARS-CoV-1. *N Engl J Med*. 2020;382(Mar (16)):1564–7.
- Esbin MN, Whitney ON, Chong S, Maurer A, Darzacq X, Tjian R. Overcoming the bottleneck to widespread testing: A rapid review of nucleic acid testing approaches for COVID-19 detection. *RNA*. 2020;(May), doi:<http://dx.doi.org/10.1261/rna.076232.120>.
- Fomsgaard AS, Rosenstjerne MW. An alternative workflow for molecular detection of SARS-CoV-2 – escape from the NA extraction kit shortage, Copenhagen, Denmark, March 2020. *Euro Surveill* 2020;25(Apr (14)):2000398.
- Geller C, Varbanov M, Duval R. Human Coronaviruses: Insights into Environmental Resistance and Its Influence on the Development of New Antiseptic Strategies. *Viruses*. 2012;4(Nov):3044–68.
- Pastorino B, Touret F, Gilles M, de Lamballerie X, Charrel R.N. Evaluation of heating and chemical protocols for inactivating SARS-CoV-2. *bioRxiv* Apr; 10.1101/2020.04.11.036855.
- Shedlovskiy D, Shcherbik N, Pestov DG. One-step hot formamide extraction of RNA from *Saccharomyces cerevisiae*. *RNA Biol*. 2017;14(Dec (12)):1722–6.
- Stead MB, Agrawal A, Bowden KE, Nasir R, Mohanty BK, Meagher RB, et al. RNA snap™: a rapid, quantitative and inexpensive, method for isolating total RNA from bacteria. *Nucleic Acids Res*. 2012;40(Jul (20)):e156–e156.0.
- Xie X, Zhong Z, Zhao W, Zheng C, Wang F, Liu J, Chest CT for Typical 2019-nCoV Pneumonia: Relationship to Negative RT-PCR Testing. *Radiology* 2020;(Feb) 200343.

Supplementary material

Materials and methods

SARS-CoV-2 standard diagnosis

Samples were collected in 2 mL volume of viral transport medium (VTM) (Deltalab) and manipulation was performed under a biosafety class II cabinet (TELSTAR bio-II-A). Diagnosis of COVID-19/SARS-CoV-2 infection was conducted with 200 μ L of VTM with standard RNA extractions, which used the MagCore viral Nucleic Acid Extraction kit (RBC Bioscience) in a MagCore liquid handler (RBC Bioscience) (lasting approximately 1 h) or the Nuclisens easyMAG kit (bioMérieux) with the eMAG liquid handler (bioMérieux) (lasting approximately 1 h and 20 min), and a one-step RT-qPCR using the LightMix® Modular SARS-CoV (COVID19) E-gene kit (TIB MOLBIOL). Alternative amplifications with primers and probes for the N and RdRp genes were available (Corman *et al.* 2020). However, given that they tended to amplify at higher cycle number, they were not used in the experiments. This commercial mix was supplemented with primers and probes for detecting the human actin (*ACTB*) gene expression to serve as an internal control of the reaction as described elsewhere (Fenollar *et al.* 2016) (Table S1). All the experiments included a non-template control and a positive control for the E-gene amplification (LightMix® kit Sarbecovirus ivRNA control +, TIB MOLBIOL). The RT-qPCR was performed in 10 μ L final volume reactions (5 μ L of sample) using a CFX96 Touch Real-Time PCR Detection System (Bio-Rad). Following the specifications, thermal cycling was performed at 55°C for 5 min for the RT step, followed by an amplification step with initial denaturation at 95°C for 5 min and 45 cycles of 95°C for 5 sec, 60°C for 15 sec, and 72°C for 15 sec. A final cooling step at 40°C for 30 sec was also included. Thermal cycling took 1 h and 14 min in total.

Alternative protocols

For the alternative protocols, 10 μ L of fresh FAE or RNAsnap™ buffers, prepared following the described conditions (Stead *et al.* 2012; Shedlovskiy *et al.* 2017) were mixed with 100 μ L of each VTM sample using a vigorous vortexing and heated at 70°C for 10 min (note that the RNAsnap™ procedure was originally described to be used at 95°C for 7 min). In parallel, a 10 μ L aliquot of each sample was also heated at 70°C for 10 min (Direct protocol). RNA integrity was not assessed before or after these treatments for safety reasons. RT-qPCR was subsequently performed with 3.5 μ L temperature-treated sample aliquots in a 10 μ L final volume using the same conditions as described for the standard diagnosis of COVID-19/SARS-CoV-2 infection. All the experiments were blind to the COVID-19 diagnosis assignment. The confusion matrix obtained in the validation stage was used to obtain sensitivity, specificity, and accuracy of the alternative sample treatment compared to the assays based on standard RNA extractions.

Statistical analysis

Data were presented as mean with standard deviation. The statistical significance for the paired differences between the means, and the sensitivity, specificity and accuracy of the protocol were assessed with MedCalc (MedCalc Software Ltd.). *P* value <0.05 was considered statistically significant.

Table S1. Oligonucleotide sequences used to detect the *ACTB* gene expression as an internal reaction control for SARS-CoV-2 detection.

Name	Nucleotide sequence (labels)
ACTB-Probe	5'-(VIC)-CGGGAAATCGTGCGTGACATTAAG-(MGB-NFQ)-3'
ACTB-F	5'-CATGCCATCCTGCGTCTGGA-3'
ACTB-R	5'-CCGTGGCCATCTCTTGCTCG-3'

Table S2. Demographic and clinical data for the 41 COVID-19 positive patients used in the validation stage.

Variable	Category	Observations
Median age, years (IQR)		72 (46-80)
Gender, male %		61.9
Comorbidities, % (n=31)*		
	Hypertension	38.7
	Diabetes	29.0
	Smoker	25.8
	Cardiomyopathy	9.7
	Oncologic	9.7
	Asthma	6.4
	Obesity	6.4
	Alcoholism	6.4
	Other systemic immune disorders	12.9
Diagnosis, % (n=31)*		
	Lower respiratory tract infection	25.8
	Bilateral pneumonia	22.6
	Common cold	2.0
Type of patient, %		
	Outpatient clinic	70.1
	Hospitalized	25.0
	Intensive Care Unit	4.9
Symptoms on admission, % (n=35)*		
	Fever	57.1
	Cough	48.6
	Dyspnea	34.3
	Diarrhea	17.1

Nausea	14.3
Lack of appetite	8.6
Sputum production	8.6
Anosmia	2.8

Biomarkers, median (IQR)

D-dimer, ng/mL (n=40)*	1033.0 (464.5-1517.5)
Lymphocyte, %	19.4 (11.7-30.5)
Lactate dehydrogenase, U/L (n=40)*	247.0 (193.5-281.0)
Ferritin, ng/mL (n=30)*	587 (229.8-953.0)
C-reactive protein (mg/dL)	8.8 (3.0-14.2)

IQR, interquartile range.

*Data available only for the indicated subset of patients.




Figure S1. RT-qPCR amplification plots for the SARS-CoV-2 E-gene for a nasopharyngeal swab VTM sample subjected to the standard RNA extraction, and the Direct, the RNAsnap™ and FAE treatments. The horizontal blue line denotes the cycle threshold. RFU: Relative fluorescence units.

References

- Corman VM, Landt O, Kaiser M, Molenkamp R, Meijer A, Chu DKW, et al. Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR. *Euro Surveill.* 2020 Jan 22;25(3):2000045.
- Fenollar F, Minodier P, Boutin A, Laporte R, Brémond V, Noël G, et al. *Tropheryma whipplei* associated with diarrhoea in young children. *Clin Microbiol Infect.* 2016;22(10):869–74.
- Shedlovskiy D, Shcherbik N, Pestov DG. One-step hot formamide extraction of RNA from *Saccharomyces cerevisiae*. *RNA Biol.* 2017 Dec 2;14(12):1722–6.
- Stead MB, Agrawal A, Bowden KE, Nasir R, Mohanty BK, Meagher RB, et al. RNA snap™: a rapid, quantitative and inexpensive, method for isolating total RNA from bacteria. *Nucleic Acids Res.* 2012 Jul 19;40(20):e156–e156.